



calibration  
( 1 : 2 )



measurement  
( 1 : 2 )



adjust gauge 0641-08



subject to technical alterations

**type:** aerosol  
**range:** 0-30 mm  
**range of indication:**  
**digital step:** 0,01 mm  
**scale interval:**  
**measuring force:** 1 +/-10,3 N  
**weight:** ca. 390 g

acceptable deviations  
VDI/VDE/DGQ 2618 Bl. 12.1/13.1)  
**max. permissible**  
**errors "G":** 0,03 mm  
**repeatability "r":** 0,02 mm  
**reference temperature:** 20° C  
**operation temperature:** 10° C - 30° C  
**storage temperature:** -10° C - 50° C

**scale:** 1:1 (1:2)  
**drawing-nr.:** IDW-AE1210-DB\_e  
**date of issue:** 25.11.2010  
**name:** J.Heuser  
  
**revision status:**  
**revision date:**